## **CTD** report

Estelle Dumont, SAMS

# 1 Shipboard CTD (SBE911)

#### 1.1 Methodology

Two CTDs were used during the cruise: both were housed in standard stainless steel frames, and were equipped with dual T and C sensors, SBE43 oxygen sensor, Chelsea Aqua 3 fluorometer, altimeter, Wetlab CStar transmissometer and Biospherical/Licor PAR/irradiance sensor.

Twenty-four 10 litres Niskin bottles were fitted on the first CTD carousel. The system was deployed from the CTD winch on the starboard side during legs 1, 3 and 4. The second CTD only carried twelve 10 litres bottles and was deployed at the stern of the ship due to ice conditions during leg 2.

The usual procedure was to first lower the CTD to around 10m deep for the pumps to switch on. The system was then brought back up to the surface before starting the cast.



Figure 1.1-1: CTD being deployed from the mid-ship gantry.

### 1.2 Data processing

The CTD data were processed according the standards described in the SAMS CTD data Processing Protocol (Dumont and Sherwin, 2008, SAMS internal report No 257), using Seabird Data Processing version 7.18b and Matlab R2007a. The processing steps were:

- Step 1(SBE Data Processing, batch processing): modules Data Conversion, Align CTD, Cell Thermal Mass, Filter, Derive, Translate and Bottle Sum.
- Step 2 (Matlab): despiking of the 24Hz data
- Step 3 (SBE Data Processing, batch processing): modules Ascii In,, Bin Average (2db-bins) and Ascii Out
- Step 4 (Matlab): plot of the data
- Step 5 (Matlab): calibration of oxygen and salinity data on both 24Hz and 2db-bin averaged datasets (post-cruise).

#### 1.2.1 Raw data processing (SBEDataProcessing)

**Data Conversion** converted raw data from engineering units to binary .cnv files and produced the .ros files. Variables exported were scan number, pump status, Julian day, pressure [db], temperature0 [ITS-90, deg C], conductivity0 [mS/cm], temperature1 [ITS-90, deg C], conductivity1 [mS/cm], oxygen [mg/I], beam attenuation [1/m], altimeter [m], fluorescence [µg/I], beam transmission [%], PAR and depth [m].

Please note:

The primary TC sensors were labelled 0, secondary 1.

The depth exported here was only for indicative purposes in the bottle files. Accurate depth calculation was performed at the Derive stage, and this first depth removed in processed files.

#### AlignCTD was then run to compensate for sensor time-lag.

The secondary conductivity was advanced by 0.073s, as recommended by Seabird.

The oxygen sensor response was advanced relative to pressure (+5s applied here). This ensures that calculations of dissolved oxygen concentration are made using measurements from the same parcel of water.

In **Cell Thermal Mass**, a recursive filter was run to remove conductivity cell thermal mass effects from the measured conductivity. The constants used were the ones given by Seabird: thermal anomaly amplitude  $\alpha$ =0.03 and thermal anomaly time constant 1/ $\beta$ =7.

**Filter** applied a low-pass filter (value of 0.15) on the pressure and depth data, which smoothed the high frequency (rapidly changing) data. To produce zero phase (no time shift), the filter was first run forward through the data and then run backward through the data. This removed any delays caused by the filter.

At the **Derive** stage, twin densities sigma-theta (kg/m3), twin salinities (psu) and depth (m) were calculated.

The data was converted from binary to ASCII format by the module **Translate**. The data had been kept in binary format up to this stage to avoid any loss in precision that could occur when converting to Ascii.

Finally, the module **BottleSum** created the ASCII bottle files (.btl) from the .ros files, for each bottle fired during a cast. These files contain mean, standard deviation, maximum and minimum values for all variables (average of 48 scans, i.e. 2s).

#### 1.2.2 Despiking (Matlab)

The pressure, oxygen, temperature (primary and secondary) and salinity (primary and secondary) data were manually despiked. Any data recorded while the pumps were not on were deleted at this stage.

Notes on the despiking:

- When a spike occurred in the pressure, temperature or salinity data, making that/those point(s) flagged as bad, the whole corresponding scan has been deleted.

- When a spike occurred in the oxygen data, making that point flagged as bad, the erroneous value was set to NaN, and other variables of the scan (i.e. temperature, salinity, etc) were kept in the dataset (if not flagged as bad themselves).

#### 1.2.3 Averaging (SBEDataProcessing)

After going through Matlab, the data files needed to be re-formatted to be recognised by SBE Data Processing. **ASCII In** added a header to the input .asc file and output a .cnv file (XXX\_2.cnv).

The module **Bin Average** averaged the 24Hz data into 2db-bins, using the downcast data only.

Ascii Out output the bin-averaged data files as ASCII (with a simplified header).

#### 1.2.4 Datafiles

The different types of files created are (example of cast no 001):

jr219\_001\_1.cnv : non despiked, non calibrated 24Hz data

jr219\_001\_2.asc : despiked, non calibrated 24Hz and 2db-bin averaged data

jr219\_001\_3.asc : despiked, calibrated 24Hz data

jr219\_001.CTD: despiked, calibrated 2db-bin averaged data (WOCE format conventions)

jr219\_001.btl : bottle data file, non calibrated

jr219\_001.hdr : header file, describing the data processing details

### 1.3 Data calibration

#### **1.3.1 Salinity calibration**

Throughout the cruise the CTD was sampled for salinity measurements, in order to calibrate the conductivity sensors. Salinity was measured using a Guildline Autosal8400, in a temperature-controlled room onboard the ship. The CTD data used for calibration comes from the .btl files (created by the Seabird software).

For the first CTD (leg 1, 3 and 4) a total of 90 salinity samples were collected and analysed, including a few duplicate samples. The Autosal and the Seabird values were in good agreement. Six data points had a difference over 0.1 psu and were not used for calibration.



Figure 1.3-1: CTD salinity calibration data and equation for leg 1, 3 and 4.

For the second CTD (leg2) 25 salinity samples were collected and analysed, including a few duplicate samples. The correlation was not as good as for the other CTD, particularly in surface waters. This is probably because the casts were done when the ship was moored to the ice floe, in highly stratified surface waters. Often, the propellers were running to clear the ice drifting at the stern, disturbing the surface layers. Five data points, showing a difference over 0.2 psu between the Autosal and the CTD readings, were removed.



**Figure 1.3-2: CTD salinity calibration data and equation for leg 2.** The final calibration equations are summarised below in Table 1.3.1-1.

CTD package	Sensor	Calibration equation	R <sup>2</sup>
CTD 1 (leg 1, 3, 4)	Primary	SAL <sub>calib</sub> = 0.9956 x SAL <sub>ctd</sub> + 0.1582	0.9993
CTD 1 (leg 1, 3, 4)	Secondary	SAL <sub>calib</sub> = 0.9957 x SAL <sub>ctd</sub> + 0.1664	0.9993
CTD 2 (leg 2)	Primary	SAL <sub>calib</sub> = 0.9454 x SAL <sub>ctd</sub> + 1.9182	0.9997
CTD 2 (leg 2)	Secondary	$SAL_{calib} = 0.9460 \times SAL_{ctd} + 1.9012$	0.9997

Table 1.3.1-1: CTD salinity calibration equations summary.

#### 1.3.2 Dissolved Oxygen calibration

For the methodology refer to section ?.? by Elena Garcia.



Figure 1.3-3: CTD O<sub>2</sub> calibration data and equation for leg 1, 3 and 4.



Figure 1.3-4: CTD O<sub>2</sub> calibration data and equation for leg 2.

To follow WOCE data format conventions, the calibrated  $O_2$  values in the final datafiles have been converted from mg/l to µmol/kg using the formula: [µmol/Kg] = (([mg/L] / 1.42903) \* 44660) / (sigma\_theta + 1000)

#### 1.3.3 Fluorometer calibration

Please note that the fluorescence data has not been calibrated in the final dataset. Users wishing to use calibrated data should refer to the chlorophyll discrete sampling carried out throughout the cruise to determine a suitable calibration equation.

### 1.4 Comments

The CTDs have generally performed well during the cruise, although a few points are worth noting:

The dual temperature and conductivity measurements were in good agreement throughout the cruise, except at the beginning of leg 2. The secondary conductivity sensor appeared to be faulty, and was replaced after a few casts. Therefore, the secondary conductivity, salinity and density for CTD013 to 017 should not be used.

The acquisition software (Seasave) installed on the ship's CTD computer was an old version, not supporting the use of the new oxygen calibration method and coefficients. The CON file used for data acquisition was therefore an inaccurate one (keeping the old oxygen calibration method). However, the data processing was performed with a more recent version of the Seabird software, and an up-to-date CON file was used to process the data.

Some errors in the calibration coefficients of the PAR sensor and fluorometer were also found in the original CON file. These were corrected shortly after the start of the cruise, and all the data were processed using the up-to-date CON file.

In the final data, the transmittance values reported were sometimes over 100% (up to 102%), indicating that the transmissometer is probably slightly out of calibration.

A misfiring issue was encountered several times on CTD1 (leg 3 and 4) for bottle 16: the acquisition software would not report the bottle as fired, although it did close correctly. The bottle firing was not recorded in the data either, shifting the bottles recordings in the .btl files (e.g. the bottle recorded as 16 was actually bottle 17, bottle 17 was bottle 18, etc). This happened for casts number 33, 34, 43, 46, 49, 53, 56, 57, 58, 59 and 80. Those .btl files have been corrected post-cruise by re-attributing the correct bottle number to the recordings. As there was no data for bottle 16 on those casts, the data from the previous or following bottle, if fired at the same depth, was used instead. If bottle 16 was the only bottle fired at a certain depth the data has been replaced by a "bad flag" value (99999).

### 1.5 T-S diagrams



Figure 1.5-1: T-S diagram of the CTD stations in leg 1 (deepest cast ).



Figure 1.5-2: T-S diagram of the CTD stations in leg 2 (top 300m of the water column).



Figure 1.5-3: T-S diagram of the CTD stations in leg 3 (full depth casts).



Figure 1.5-4: -S diagram of the CTD stations in leg 4 (full depth casts).

The water masses represented on the diagrams are: AtW = Atlantic Water; ArW = Arctic Water; SW = Surface Water; LW = Local water; WCW= Winter Cooled Water IW = Intermediate Water; TAtW = TYransforemed Atlantic Water.

### 1.6 CTD Transects



Figure 1.6-1: Location of the CTD transects realised during leg 4.

### 1.6.1 Amsterdamøya (14/07/10 – 15/07/10)

Data (left to right) from casts 64, 65, 66, 76, 68, 69, 70 and 71.



Figure 1.6.1-1: Amsterdamoya transect temperature, for the whole depth and close-up on the 200m surface layer.



Figure 1.6-2: Amsterdamoya transect salinity, for the whole depth and close-up on the 200m surface layer.



Figure 1.6-3: Amsterdamoya transect sigma-theta, for the whole depth and close-up on the 200m surface layer.



Figure 1.6-4: Amsterdamoya transect dissolved oxygen, for the whole depth and close-up on the 200m surface layer.



Figure 1.6-5: Amsterdamoya transect fluorescence, for the whole depth and close-up on the 200m surface layer.

### 1.6.2 Norske Banken (17/07/10 - 18/07/10)

Data (left to right) from casts 81, 82, 83, 84, 85, 86, 87, 88 and 89.



Figure 1.6.2-1: Norske Banken transect temperature, for the whole depth and close-up on the 200m surface layer.



surface layer.



Figure 1.6.2-3: Norske Banken transect sigma-theta, for the whole depth and close-up on the 200m surface layer.



Figure 1.6.2-4: Norske Banken transect dissolved oxygen, for the whole depth and close-up on the 200m surface layer.



Figure 1.6.2-5: Norske Banken transect fluorescence, for the whole depth and close-up on the 200m surface layer.

### 1.6.3 Isfjord Banken (20/07/10)

Data (left to right) from casts 100, 99, 98, 97 and 96.



Figure 1.6.3-1: Isfjord Banken transect temperature, for the whole depth and close-up on the 200m surface layer.



Figure 1.6.3-2: Isfjord Banken transect salinity, for the whole depth and close-up on the 200m surface layer.



Figure 1.6.3-3: Isfjord Banken transect sigma-theta, for the whole depth and close-up on the 200m surface layer.



Figure 1.6.3-4: Isfjord Banken transect dissolved oxygen, for the whole depth and close-up on the 200m surface layer.



Figure 1.6.3-5: Isfjord Banken transect fluorescence, for the whole depth and close-up on the 200m surface layer.

#### 1.7 Instrument specifications

#### 1.7.1 SBE911 Leg 1, 3, 4

Configuration report for SBE 911plus/917plus CTD

```
Frequency channels suppressed : 0HVoltage words suppressed : 0IComputer interface : RS-232CJScans to average : 1CTcorNMEA position data added : NoCPcorNMEA depth data added : NoSlopeNMEA time added : NoOffsetSurface DPB woltage added : NoVo
                                                                   : 1.55766840e+000
: -1.88282087e-003
: 2.27423366e-004

      NMEA position data added
      : 1

      NMEA depth data added
      : No

      NMEA time added
      : Wo

                                                                        : 3.2500e-006
                                                                         : -9.57000000e-008
                                                                       : 1.00000000
                                : No
: No
: No
                                                                        : 0.00000
Surface PAR voltage added
                                                  6) A/D voltage 0, PAR/Irradiance,
Scan time added
                                                     Biospherical/Licor
                                                                                 : 7274
1) Frequency 0, Temperature
                                                         Serial number
                                                                                : 12/1/2009
    Serial number : 5042
                                                         Calibrated on
    Calibrated on : 12/04/2008
                                                        М
                                                                                 : 1.00000000
           : 4.33201372e-003
                                                       В
                                                                                 : 0.00000000
   G
                   : 6.33549986e-004
                                                        Calibration constant :
   н
   I
                   : 2.08531346e-005
                                                     55248618784.50000000
   т
Ј
                  : 1.84440097e-006
                                                      Multiplier
                                                                                 : 1.00000000
   F0 : 1000.000
Slope : 1.0000000
Offset : 0.0000
                                                        Offset
                                                                                : -0.02596740
                                                     7) A/D voltage 1, Free
2) Frequency 1, Conductivity
                                                     8) A/D voltage 2, Oxygen, SBE 43
    Serial number : 3488
                                                         Serial number : 0245
   Calibrated on : 22/04/2008
                                                        Calibrated on : 10/12/2008
         : -1.02006618e+001
: 1.56812636e+000
    G
                                                       Equation : Sea-Bird
                                                                         : 3.97100e-001
   Н
                                                        Soc
   I
J
                                                      Soc
Offset
                                                                     : -4.31400e-001
: -4.94730e-004
: 1.36160e-004
                  : -2.01455756e-003
                 : 2.36866095e-004
: 3.2500e-006
: -9.57000000e-008
                                                      A
B
   CTcor
   CPcor : 3.2500e-006
CPcor : -9.57000000
Slope : 1.000000
Offset : 0.00000
                                                      С
                                                                        : -2.43420e-006
                                                                        : 3.60000e-002
: 1.20000e+000
                                                        E
                                                       Tau20
                                                       D1
                                                                        : 1.92630e-004
3) Frequency 2, Pressure, Digiquartz with
                                                        D2
                                                                         : -4.64800e-002
TС
                                                       Н1
                                                                        : -3.30000e-002
                                                        H2
••• 3
   Serial number : 0541-75429
                                                                         : 5.00000e+003
   Calibrated on : 18/07/2007
                                                                         : 1.45000e+003
           : -4.398881e+004
   C1
                 : -4.398881e+004

: -5.551403e-001

: 1.279490e-002

: 3.603000e-002

: 0.000000e+000

: 2.986716e+001

: -5.274889e-004

: 4.092900e-006

: 4165000
   C2
                                                    9) A/D voltage 3, Altimeter
   C3
                                                       Serial number : 2130.27001
   D1
                                                         Calibrated on : no calibration
   D2
                                                     available
   т1
                                                      Scale factor : 15.000
   т2
                                                                         : 0.000
                                                        Offset
   ΤЗ
   т4
                  : 1.616590e-009
: 0.000000e+000
                                                     10) A/D voltage 4, Fluorometer, Chelsea
   т5
                                                     Aqua 3
                 : 0.99994000
: 0.52570
: 1.287420e-002
                                                         Serial number : 088-249
   Slope
   Offset
AD590M
                                                         Calibrated on : 27/08/2009
                                                         VB : 0.154130
                  : -8.793390e+000
   AD590B
                                                         V1
                                                                          : 2.015700
                                                          Vacetone
                                                                          : 0.199600
                                                          Scale factor : 1.000000
4) Frequency 3, Temperature, 2
                                                          Slope
   Serial number : 5043
                                                                          : 1,000000
    Calibrated on : 09/04/2008
                                                          Offset
                                                                          : 0.000000
    G : 4.34448076e-003
                    : 6.34699898e-004
                                                     11) A/D voltage 5, Free
   Η
                 : 2.11669515e-005
: 1.90896423e-006
: 1000.000
: 1.00000000
   Т
   ъ
                                                     12) A/D voltage 6, Transmissometer,
   FO
                                                     Chelsea/Seatech/Wetlab CStar
             : 1.00000
: 0.0000
   Slope
                                                        Serial number : CST-396DR
                                                          Calibrated on : 23/08/2007
   Offset
                                                        M
                                                                         : 22,2100
5) Frequency 4, Conductivity, 2
                                                       В
                                                                          : -1.3104
    Serial number : 3491
                                                        Path length : 0.250
    Calibrated on : 22/04/2008
                  : -1.01164027e+001 13) A/D voltage 7, Free
   G
```

#### 1.7.2 SBE911 Leg 2

Configuration report for SBE 911plus/917plus CTD

```
Frequency channels suppressed : 0
Voltage words suppressed : 0
Computer interface
                             : 1
: No
Scans to average
NMEA position data added : No
NMEA depth data added : No
NMEA time added
                              : No
Surface PAR voltage added
                              : No
                             : No
Scan time added
1) Frequency 0, Temperature
   Serial number : 2307
   Calibrated on : 20/7/2007
        : 4.33439937e-003
   G
                 : 6.44608874e-004
  Н
                : 2.37793994e-005
  Т
                : 2.31412704e-006
: 1000.000
   ъ
   FO
               : 1.00000000
   Slope
                : 0.0000
   Offset
2) Frequency 1, Conductivity
   Serial number : 1913
   Calibrated on : 17/7/2007
         : -4.02881844e+000
: 5.32503936e-001
   G
  Н
                : -6.89373545e-004
   I
               : 6.23129385e-005
: 3.2500e-006
   J
   CTcor
               : -9.57000000e-008
: 1.00000000
: 0.00000
   CPcor
   Slope
   Offset
3) Frequency 2, Pressure, Digiquartz with
тС
   Serial number : 09p30856-0707
   Calibrated on : 13/06/2007
           : -4.925971e+004
: -2.136250e-001
   C1
   C2
               : 9.435710e-003
: 3.900400e-002
: 0.000000e+000
   C3
   D1
  D2
               : 2.983458e+001
: -3.883229e-004
   т1
   т2
                : 3.262440e-006
   т3
                : 3.429810e-009
: 0.000000e+000
   т4
   Τ5
               : 1.00005000
: -0.81450
   Slope
   Offset
               : 1.277500e-002
   AD590M
   AD590B
                 : -9.391460e+000
4) Frequency 3, Temperature, 2
   Serial number : 4472
   Calibrated on : 11/7/2007
        : 4.41422567e-003
   G
  Н
                 : 6.43319527e-004
   I
                : 2.23249357e-005
                : 1.96317455e-006
: 1000.000
   J
   F0
   Slope
                : 1.00000000
```

```
Offset
                              : 0.0000
: RS-232C 5) Frequency 4, Conductivity, 2
              Serial number : 2255
                Calibrated on : 17/7/2007
               G : -1.02637082e+001
              H
I
J
CTcor
                             : 1.41424921e+000
: -2.97291304e-003
                             : 3.13705184e-004
                              : 3.2500e-006
               CPcor
                             : -9.57000000e-008
                             : 1.00000000
: 0.00000
                 Slope
                 Offset
              6) A/D voltage 0, PAR/Irradiance,
              Biospherical/Licor
                Serial number
                                    : 7275
                 Calibrated on
                                    : 26/7/2007
                М
                                    : 1.00000000
                В
                                     : 0.00000000
                Calibration constant :
              59523809523.80999800
                Multiplier
                                     : 1.00000000
                                    : -0.02516571
                Offset
              7) A/D voltage 1, Free
              8) A/D voltage 2, Oxygen, SBE 43
                Serial number : 0242
                 Calibrated on : 12/06/2007
                 Equation : Owens-Millard
Soc : 3.8420e-001
                Soc
                            : 0.0000
: -0.4887
                Boc
                Offset
                 Tcor
                             : 0.0003
: 1.35e-004
                 Pcor
                             : 0.0
                Tau
              9) A/D voltage 3, Altimeter
                 Serial number : 2130.27001
                 Calibrated on : no calibration
              available
                 Scale factor : 15.000
                 Offset
                              : 0.000
              10) A/D voltage 4, Fluorometer, Chelsea
              Aqua 3
                  Serial number : 088-249
                  Calibrated on : 13/09/2007
                 VB : 0.181000
                 V1
                               : 2.097600
                             : 0.202800
                 Vacetone
                  Scale factor : 1.000000
                  Slope
                               : 1.000000
                 Offset
                               : 0.000000
             11) A/D voltage 5, Free
             12) A/D voltage 6, Free
```

13) A/D voltage 7, Free

# 2 Hand-help CTD (SBE19+)

### 2.1 Methodology

This CTD was used during leg 2 for a) profiling in the water sampling hole drilled in the ice; and b) in association with under-ice light sensors for profiles and overnight deployments from the diving hole (see details in the PAR/ Under-ice light measurements section). It was also deployed once from the ship on leg 3 in association with the light spectrometer and SAtlantics sensors.

All data was recorded at 4Hz (maximum sampling rate).



Figure 2.1-1: SBE19+ and SAtlantics light sensors installed on the board and profiling under ice.

### 2.2 Data processing

The CTD data were processed according the recommended Seabird procedures, using Seabird Data Processing version 7.18b. The processing steps were: Data Conversion, Filter, Align CTD, Cell Thermal Mass, Derive, Translate, Bin Average (1 second) and Ascii Out.

### 2.3 Data summary

Data file	Date	Start	Event	Description
		Time	No	
sbe19+_5215_260610_a	26-Jun-10	08:13	l18	Cast associated with water sampling
sbe19+_5215_260610_b	26-Jun-10	18:07	125	Under-ice CTD/light profile (test)
sbe19+_5215_270610	27-Jun-10	11:41	131	Under-ice CTD/light profile
sbe19+_5215_280610	28-Jun-10	12:41	136	Under-ice CTD/light profile
sbe19+_5215_290610_a	29-Jun-10	10:00	l41	Cast associated with water sampling
sbe19+_5215_290610_b	29-Jun-10	11:52	144	Cast associated with water sampling
sbe10 + 5215 200610 c	29-Jun-10	17:34	145	Under-ice CTD/light overnight
sbe19+_3213_290010_c				deployment
sbe19+_5215_300610_a	30-Jun-10	08:48	147	Under-ice CTD/light profiles
sbe19+_5215_300610_b	30-Jun-10	15:05	155	Under-ice CTD/light profiles
sbe19+_5215_300610_c	30-Jun-10	19:35	156	Under-ice CTD/light profile + overnight
				deployment
sbe19+_5215_100710	10-Jul-10	10:02	145	CTD/light profile from ship

### 2.4 Instrument specifications

Configuration report for SBE 19plus Seacat CTD Pressure sensor type : Strain Gauge External voltage channels : 4 : Profile Mode Scans to average : 1 NMEA position data added : No NMEA depth data added : No NMEA time added : No Surface PAR voltage added : No Scan time added : No 1) Count, Temperature Serial number : 5215 Calibrated on : 07-Aug-07 : 1.20495500e-003 : 2.77321400e-004 A0 A1 A2 : -1.96517300e-006 A3 : 2.21643700e-007 Slope : 1.0000000 Offset : 0.0000 2) Frequency 0, Conductivity Serial number : 5215 Calibrated on : 07-Aug-07 : -1.02557600e+000 : 1.57749000e-001 G Η : -5.98424400e-004 : 7.12694500e-005 : 3.2500e-006 I J CTcor CPcor : -9.57000000e-008 : 1.00000000 Slope Offset : 0.00000 3) Count, Pressure, Strain Gauge Serial number : 5215 Calibrated on : 01-Aug-07 PA0 : 1.34278800e+000 PA1 : 2.63258600e-003 : 2.63258600e-003 : 1.71027200e-011 : -6.22525800e+001 : 5.25907100e+001 : -2.89748100e-001 : 5.24870500e+005 : 3.67886900e+001 PA2 PTEMPA0 PTEMPA1 PTEMPA2 PTCA0 PTCA1 : 3.6/888300000001 : -6.450026000001 : 2.5604620000000 : -4.750000000000000 : 0.0000000000000 : 0.0000000 PTCA2 PTCB0 PIC PTCB0 PTCB1 TCB2 4) A/D voltage 0, Fluorometer, Wetlab Wetstar Serial number : 1359 Calibrated on : 10/05/2010 Vblank : 0.071 Scale factor : 15.000 5) A/D voltage 1, Free 6) A/D voltage 2, Free 7) A/D voltage 3, Free