



Sea-Bird Electronics, Inc.

1808 136th Place NE, Bellevue, Washington 98005 USA
Website: <http://www.seabird.com>

Phone: (425) 643-9866
FAX: (425) 643-9954
Email: seabird@seabird.com

SBE Pressure Test Certificate

Test Date: 12/30/2004 Description SBE-9Plus CTD

Job Number: 37898 Customer Name SOC

SBE Sensor Information:

Model Number: 09
Serial Number: 0782

Pressure Sensor Information:

Sensor Type: DigiQuartz
Sensor Serial Number: 94756
Sensor Rating: 10000

Pressure Test Protocol:

Low Pressure Test: 50 PSI Held For 15 Minutes

High Pressure Test: 10000 PSI Held For 15 Minutes

Passed Test:

Tested By: pc

