

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2165
CALIBRATION DATE: 29-Apr-05

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.07746602e+001
h = 1.48155068e+000
i = -2.48043447e-003
j = 2.47324677e-004
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 3.92089824e-008
b = 1.47447594e+000
c = -1.07586702e+001
d = -8.23467477e-005
m = 7.4
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.70123	0.00000	0.00000
-0.9999	34.9335	2.81314	5.13525	2.81312	-0.00003
1.0269	34.9338	2.98739	5.24902	2.98742	0.00004
15.0001	34.9344	4.28457	6.02825	4.28457	0.00000
18.5001	34.9345	4.63236	6.22047	4.63233	-0.00004
29.0001	34.9325	5.71921	6.78592	5.71928	0.00006
32.5001	34.9275	6.09319	6.96970	6.09315	-0.00004

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

